

\* =mandatory field)

- **Investigator:\***
    - **Name\*:** Dr. Christopher Sabine
    - **Organization:** NOAA/PMEL
    - **Address:**  
NOAA/PMEL  
7600 Sand Point Way NE  
Seattle Washington, 98115 USA  
**Phone:** 206-526-4809
    - **Email:** [chris.sabine@noaa.gov](mailto:chris.sabine@noaa.gov)
  - **Dataset Info:\***
    - Dataset\_ID\*: [MOSEAN\\_158W\\_23N\\_May2005\\_Jan2006](#)
    - **Submission Dates:\***
      - Initial\_Submission: [20110217](#) (YYYYMMDD)
      - Revised\_Submission: (YYYYMMDD)
  - **Cruise Info:\***
    - **Experiment:**
      - Experiment\_Name\*:
      - **Cruise:( - )**
        - Cruise\_ID: (EXPOCODE)
        - Section: (Leg)
        - **Geographical Coverage:\***
          - Geographical\_Region:
          - **Bounds:**
            - Westernmost\_Longitude:  
Enter decimal fractions of degrees:  
or Degrees, Minutes, Seconds:
            - Easternmost\_Longitude:  
Enter decimal fractions of degrees: [- 158.11](#) (+ = E, - = W)  
or Degrees, Minutes, Seconds:
            - Northernmost\_Latitude:  
Enter decimal fractions of degrees:[+22.76](#) (+ = E, - = W)
            - Southernmost\_Latitude:  
Enter decimal fractions of degrees:
      - **Temporal Coverage:**
        - Start\_Date: [20050529](#) (YYYYMMDD)
        - End\_Date: [20060120](#) (YYYYMMDD)
  - **Vessel:\*** [Mooring platform](#)
    - Vessel\_Name:
    - Vessel\_ID:
    - Country:
    - Vessel\_Owner:
- **Variables Info:\***
  - **Variable:**
    - Variable\_Name and Description\*:
- [xCO<sub>2</sub> SW \(wet\) \(umol/mol\) - Mole fraction of CO<sub>2</sub> in air in equilibrium with the seawater at sea surface temperature and measured humidity.](#)
- [CO<sub>2</sub> SW QF – Quality Flag for xCO<sub>2</sub> SW \(wet\).](#)
- [H<sub>2</sub>O SW \(mmol/mol\) - Mole fraction of H<sub>2</sub>O in air from equilibrator .](#)
- [xCO<sub>2</sub> Air \(wet\) \(umol/mol\) - Mole fraction of CO<sub>2</sub> in air from airblock, 4 feet above the sea surface at measured humidity.](#)
- [CO<sub>2</sub> Air QF – Quality Flag for xCO<sub>2</sub> Air \(wet\)](#)
- [H<sub>2</sub>O Air \(mmol/mol\) - Mole fraction of H<sub>2</sub>O in air from airblock, 4 feet above the sea surface.](#)

- Licor Atm Pressure (hPa) – Atmospheric pressure at the airblock, 4 feet above the sea surface
- Licor Temp (C) – Temperature of the Infrared Licor 820 in degrees Celsius
- % O<sub>2</sub> - The percent oxygen of the surface seawater divided by the percent oxygen of the atmosphere at 4 feet above the sea surface. Disclaimer: The oxygen measurement is made in the equilibrated air. We have found that the oxygen does not come to complete equilibrium so any rapid changes in oxygen do not get properly captured using this system. Therefore, we tend to use the oxygen data only as a qualitative sense of the biology. It is not a quantitative measure.
- SST (C) - Sea Surface Temperature collected by UCSB/OPL. UCSB/OPL provide internally recorded temperature data at 2 minute resolution. The temperature collected during the equilibration period is reported in this dataset. UCSB/OPL advises to check the UCSB site at the time of use for the most accurate data available.
- Salinity - Salinity at 75m depth is collected by UCSB/OPL. UCSB/OPL provide internally recorded salinity data at 2 minute resolution. The salinity collected during the equilibration period is reported in this dataset. UCSB/OPL advises to check the UCSB site at the time of use for the most accurate data available.

There was no sea surface salinity available during this deployment; however, analysis of bottle salinity measurements at the surface collected by the Hawaii Ocean Time Series Project compared to the 75m data reveal that the largest salinity difference (0.4 psu) found between depths does not change calculated fCO<sub>2</sub>.

- xCO<sub>2</sub> SW (dry) (umol/mol) – Mole fraction of CO<sub>2</sub> in air in equilibrium with the seawater at sea surface temperature (dry air).
- xCO<sub>2</sub> Air (dry) (umol/mol) – Mole fraction of CO<sub>2</sub> in air at the airblock, 4 feet above the sea surface (dry air).
- fCO<sub>2</sub> SW (sat) uatm – Fugacity of CO<sub>2</sub> in air in equilibrium with the seawater at sea surface temperature (100% humidity). Since the measurements are taken at the sea surface, warming calculations are not necessary.
- fCO<sub>2</sub> Air (sat) uatm – Fugacity of CO<sub>2</sub> in air at the airblock, 4 feet above the sea surface (100% humidity).
- dfCO<sub>2</sub> – Difference of the fugacity of the CO<sub>2</sub> in seawater and the fugacity of the CO<sub>2</sub> in air (fCO<sub>2</sub> SW - fCO<sub>2</sub> Air).

- **Method\_Description:\***

- **Equilibrator\_Design:**

- Equilibrator\_Type: (show pick list) Bubble Equilibrator
    - Equilibrator\_Volume: (L) N/A
    - Water\_Flow\_Rate: (L/min) N/A
    - Headspace\_Gas\_Flow\_Rate: (L/min) ~600 cc/min
    - Vented: (show pick list) Yes

- Measurement\_Method: Absolute, non-dispersive infrared (NDIR) gas analyzer

- Manufacturer\_of\_Calibration\_Gas: NOAA Earth System Research Laboratory (ESRL)

- **CO<sub>2</sub>\_Sensors:**

- **CO<sub>2</sub>\_Sensor:**

- Manufacturer: Licor
      - Model: Environmental\_Control: LI-820
      - Resolution: 0.01 ppm
      - Uncertainty: < 2.5% of reading with 14 cm bench (stated)  
<1.5 ppm determined in lab
    - CO<sub>2</sub>\_Sensor\_Calibration: (For each calibration gas, document traceability to an internationally recognized scale, including date and place of last calibration. Include uncertainty of assigned value.)

At the beginning of each sample, the instrument self-calibrates using a zero and high standard. The zero standard is generated by cycling a small amount of air through a soda lime chamber. The high standard is from a cylinder of calibrated standard reference gas, 503.3  $\mu\text{mol/mol}$  from ESRL. ESRL standards are traceable to WMO x93 scale with a stated reproducibility of 0.06 micromole/mole.

- **Other\_Sensors:**
  - Manufacturer: Maxtec
  - Model: Max-250
  - Resolution: 0.01 %
  - Uncertainty:  $\pm 2.0\%$  Full Scale over operating temperature range
  - Calibration: (For each sensor of pressure, temperature, and salinity, document traceability to an internationally recognized scale, including date and place of last calibration.)  
Factory calibrated before purchase. Recalibrated to sea level atmospheric air every 7 days.
- **Other\_Sensors:**
  - Manufacturer: Sensirion
  - Model: SHT71
  - Resolution: 0.01 %
  - Uncertainty: Measurement range: 0-100% RH  
Absolute. RH accuracy:  $\pm 3\%$  RH (20-80% RH)  
Repeatability RH:  $\pm 0.1\%$  RH
  - Calibration: (For each sensor of pressure, temperature, and salinity, document traceability to an internationally recognized scale, including date and place of last calibration.)  
Factory calibrated before purchase.
- Method\_References: (Publication(s) describing method)

Sabine, C. (2005): High-resolution ocean and atmosphere  $\text{pCO}_2$  time-series measurements. The State of the Ocean and the Ocean Observing System for Climate, Annual Report, Fiscal Year 2004, NOAA/OGP/Office of Climate Observation, Section 3.32a, 246–253.

- Additional Information

- All measurements are at sea surface temperature and atmospheric pressure.
- During the equilibration cycle, a closed loop of air equilibrates with seawater for 10 minutes. Once the equilibration period is complete, the pump stops and the system opens to the atmosphere allowing the pressure to equilibrate with atmospheric pressure. Measurements are recorded for 30 seconds at 2 hertz and then averaged.
- During the air cycle, fresh air is pumped through the detector for 1 minute. Once the pump stops, the system opens to the atmosphere allowing the pressure to equilibrate with atmospheric pressure. Measurements are recorded for 30 seconds at 2 hertz and then averaged.
- The gas streams for both the air cycle and equilibrator cycle are partially dried before entering the detector. The values listed as wet  $\text{xCO}_2$  generally have relative humidity levels ranging from 40 to 80 percent. The humidity levels increase over the course of a deployment.
- Sampling occurs every 3 hours. The infrared detector is calibrated at the beginning of every sampling period. Averaged data and standard deviations for each measurement are transmitted back daily.

- To calculate the dry measurements, the water mole fraction in the Licor detector must be known. A relative humidity sensor is located immediately downstream of the detector.

- As part of the QC process, each data set is compared with the Marine Boundary Layer (MBL) data from GlobalView-CO<sub>2</sub>. The data from this deployment were  $0.1 \pm 0.7$  umol/mol on average of the MBL data and therefore no correction was applied.

GLOBALVIEW-CO<sub>2</sub>: Cooperative Atmospheric Data Integration Project - Carbon Dioxide. CD-ROM, NOAA ESRL, Boulder, Colorado [Also available on Internet via anonymous FTP to ftp.cmdl.noaa.gov, Path: ccg/co2/GLOBALVIEW], 2010

-During the QC process, an adjustment to the Licor pressure is typically made based on each sensor's bias to barometric pressure as measured in the lab. We have not yet run this test on this system.

- No data = -9.999 or -999

- Data\_set\_References: (Publication(s) describing data set) **None**
- Citation: (How to cite this data set) **Sabine, C., S. Maenner and A. Sutton. 2011. High-resolution ocean and atmosphere pCO<sub>2</sub> time-series measurements from mooring MOSEAN.**
- **Data\_Set\_Link:**
  - URL\*: <http://www.pmel.noaa.gov/co2/story/HALE-ALOHA>
- Label\*: **PMEL CO<sub>2</sub> Group – MOSEAN mooring** 
  - Link\_Note: (Optional instructions or remarks)(**m s t**)

Quality Flags definitions:

- 2 = Acceptable measurement;
- 3 = Questionable measurement;
- 4 = Bad measurement
- 5 = Not reported;
- 9 = Sample not drawn for this measurement from this bottle.

#### Quality Flag Log for this dataset.

Date	Measurement	Value (Dry)	Flag	Comments
6/6/2005 15:16	xCO <sub>2</sub> _SW	370.0476196	3	CO2 data submitted was adjusted by + 4 ppm b/c span calibration was off as predicted by change in Licor temperature
6/6/2005 15:16	xCO <sub>2</sub> _Air	382.9005694	3	CO2 data submitted was adjusted by + 4 ppm b/c span calibration was off as predicted by change in Licor temperature
6/13/2005 6:16	xCO <sub>2</sub> _Air	381.5297784	4	high CO2 air st dev
6/14/2005 12:16	xCO <sub>2</sub> _SW	374.2042583	3	CO2 data submitted was adjusted by + 4 ppm b/c span calibration was off as predicted by change in Licor temperature
6/14/2005 12:16	xCO <sub>2</sub> _Air	382.2121867	3	CO2 data submitted was adjusted by + 4 ppm b/c span calibration was off as predicted by change in Licor temperature
6/16/2005 21:16	xCO <sub>2</sub> _Air	386.0563374	4	high CO2 air st dev and double check Span flag after
6/18/2005 21:16	xCO <sub>2</sub> _SW	365.088491	3	CO2 data submitted was adjusted by + 4 ppm b/c span calibration was off as predicted by change in Licor temperature
6/18/2005 21:16	xCO <sub>2</sub> _Air	378.092321	3	CO2 data submitted was adjusted by + 4 ppm b/c span calibration was off as predicted by change in Licor temperature

6/20/2005 3:16 xCO2\_SW 379.1976495 3 CO2 data submitted was adjusted by - 4 ppm b/c span calibration was off as predicted by change in Licor temperature

6/20/2005 3:16 xCO2\_Air 386.1970896 3 CO2 data submitted was adjusted by - 4 ppm b/c span calibration was off as predicted by change in Licor temperature

6/29/2005 12:16 xCO2\_SW 378.6254493 3 CO2 data submitted was adjusted by - 4 ppm b/c span calibration was off as predicted by change in Licor temperature

6/29/2005 12:16 xCO2\_Air 385.3187225 3 CO2 data submitted was adjusted by - 4 ppm b/c span calibration was off as predicted by change in Licor temperature

7/8/2005 6:16 xCO2\_SW 371.9365235 3 CO2 data submitted was adjusted by + 5 ppm b/c span calibration was off as predicted by change in Licor temperature

7/8/2005 6:16 xCO2\_Air 375.039343 3 CO2 data submitted was adjusted by + 5 ppm b/c span calibration was off as predicted by change in Licor temperature

7/8/2005 21:16 xCO2\_Air 379.4373672 4 high CO2 air st dev

7/13/2005 12:16 xCO2\_SW 370.524585 3 CO2 data submitted was adjusted by + 4 ppm b/c span calibration was off as predicted by change in Licor temperature

7/13/2005 12:16 xCO2\_Air 375.7450998 3 CO2 data submitted was adjusted by + 4 ppm b/c span calibration was off as predicted by change in Licor temperature

7/26/2005 3:16 xCO2\_Air 376.3453581 4 high CO2 air st dev

8/13/2005 18:16 xCO2\_Air 375.4055509 4 high CO2 air st dev

9/4/2005 3:16 xCO2\_SW 385.1451789 3 CO2 data submitted was adjusted by + 3 ppm b/c span calibration was off as predicted by change in Licor temperature

9/4/2005 3:16 xCO2\_Air 371.2530952 3 CO2 data submitted was adjusted by + 3 ppm b/c span calibration was off as predicted by change in Licor temperature

9/8/2005 9:16 xCO2\_Air 374.3930989 4 high CO2 air st dev

9/28/2005 12:16 xCO2\_SW 377.8608986 3 CO2 data submitted was adjusted by + 4 ppm b/c span calibration was off as predicted by change in Licor temperature

9/28/2005 12:16 xCO2\_Air 374.9916702 3 CO2 data submitted was adjusted by + 4 ppm b/c span calibration was off as predicted by change in Licor temperature

10/15/2005 15:16 xCO2\_SW 374.8960924 3 CO2 data submitted was adjusted by - 5 ppm b/c span calibration was off as predicted by change in Licor temperature

10/15/2005 15:16 xCO2\_Air 376.477895 3 CO2 data submitted was adjusted by - 5 ppm b/c span calibration was off as predicted by change in Licor temperature

10/19/2005 9:16 xCO2\_Air 376.3223681 4 high CO2 air st dev

10/21/2005 6:16 xCO2\_Air 375.8482689 4 high CO2 air st dev

10/28/2005 18:16 xCO2\_SW 375.0283532 3 CO2 data submitted was adjusted by - 4 ppm b/c span calibration was off as predicted by change in Licor temperature

10/28/2005 18:16 xCO2\_Air 378.5169082 3 CO2 data submitted was adjusted by - 4 ppm b/c span calibration was off as predicted by change in Licor temperature

10/29/2005 9:16 xCO2\_SW 379.0163451 3 CO2 data submitted was adjusted by - 5 ppm b/c span calibration was off as predicted by change in Licor temperature

10/29/2005 9:16 xCO2\_Air 382.6804279 3 CO2 data submitted was adjusted by - 5 ppm b/c span calibration was off as predicted by change in Licor temperature

10/30/2005 12:16 xCO2\_SW 379.2360831 3 CO2 data submitted was adjusted by + 4 ppm b/c span calibration was off as predicted by change in Licor temperature

10/30/2005 12:16 xCO2\_Air 378.407328 3 CO2 data submitted was adjusted by + 4 ppm b/c span calibration was off as predicted by change in Licor temperature

11/13/2005 9:16 xCO2\_SW 384.3704398 3 CO2 data submitted was adjusted by - 4 ppm b/c span calibration was off as predicted by change in Licor temperature

11/13/2005 9:16 xCO2\_Air 381.8168036 3 CO2 data submitted was adjusted by - 4 ppm b/c span calibration was off as predicted by change in Licor temperature

11/27/2005 18:16 xCO2\_Air 390.223949 4 bad CO2 air measurment - likely pressure problem

12/6/2005 9:16 xCO2\_SW 380.4907998 3 CO2 data submitted was adjusted by - 4 ppm b/c span calibration was off as predicted by change in Licor temperature

12/6/2005 9:16 xCO2\_Air 382.8919917 3 CO2 data submitted was adjusted by - 4 ppm b/c span calibration was off as predicted by change in Licor temperature

12/8/2005 21:16 xCO2\_SW 381.3224822 3 CO2 data submitted was adjusted by + 4 ppm b/c span calibration was off as predicted by change in Licor temperature

12/8/2005 21:16	xCO2_Air	381.6918305	3	CO2 data submitted was adjusted by + 4 ppm b/c
span calibration was off as predicted by change in Licor temperature				
12/10/2005 0:16	xCO2_SW	381.8447551	3	CO2 data submitted was adjusted by - 4 ppm b/c
span calibration was off as predicted by change in Licor temperature				
12/10/2005 0:16	xCO2_Air	381.6552503	3	CO2 data submitted was adjusted by - 4 ppm b/c
span calibration was off as predicted by change in Licor temperature				
1/1/2006 21:16	xCO2_SW	369.9877456	4	high CO2 sw st dev
1/2/2006 6:16	xCO2_SW	370.423058	4	high CO2 sw st dev
1/4/2006 3:16	xCO2_SW	376.021908	3	CO2 data submitted was adjusted by + 4 ppm b/c span
calibration was off as predicted by change in Licor temperature				
1/4/2006 3:16	xCO2_Air	382.2739343	3	CO2 data submitted was adjusted by + 4 ppm b/c span
calibration was off as predicted by change in Licor temperature				
1/7/2006 6:16	xCO2_Air	378.2593329	4	high CO2 air st dev
1/17/2006 6:16	xCO2_SW	358.7570933	4	high CO2 sw st dev